



# Standard Test Method for Collection and Analysis of Hexavalent Chromium in Ambient Atmospheres<sup>1</sup>

This standard is issued under the fixed designation D 5281; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This test method covers the collection and measurement of hexavalent chromium [Cr(VI)] in the ambient atmosphere.

1.2 This test method collects and stabilizes atmospheric hexavalent chromium using an alkaline impinger buffer solution in a wet impingement sampling technique. Lead chromate [PbCrO<sub>4</sub>], generally considered poorly soluble in water, is soluble in the impinger solution up to 940  $\mu\text{g/L}$  as hexavalent chromium.

1.3 This test method measures hexavalent chromium using an ion chromatographic separation combined with a post separation reaction with a colorimetric reagent and photometric detection.

1.4 This test method is applicable in the range from 0.2 to 100  $\text{ng/m}^3$  of hexavalent chromium in the atmosphere assuming 20  $\text{m}^3$  of air sample. The range can be extended upwards by appropriate dilution.

1.5 The values stated in SI units are to be regarded as the standard. The inch-pound units given in parentheses are for information only.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

## 2. Referenced Documents

### 2.1 ASTM Standards:

D 1193 Specification for Reagent Water<sup>2</sup>

D 1356 Terminology Relating to Sampling and Analysis of Atmospheres<sup>3</sup>

D 1357 Practice for Planning the Sampling of the Ambient Atmosphere<sup>3</sup>

D 2914 Test Methods for Sulfur Dioxide Content of the Atmosphere (West-Gaeke Method)<sup>3</sup>

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D22 on Sampling and Analysis of Atmospheres and is the direct responsibility of Subcommittee D22.03 on Ambient Atmospheres and Source Emissions.

Current edition approved October 10, 1998. Published December 1998. Originally published as D 5281 – 92. Last previous edition D 5281 – 92 (1997) <sup>$\epsilon$ 1</sup>.

<sup>2</sup> Annual Book of ASTM Standards, Vol 11.01.

<sup>3</sup> Annual Book of ASTM Standards, Vol 11.03.

D 3195 Practice for Rotameter Calibration<sup>3</sup>

D 3586 Test Method for Chromium in Workplace Atmospheres (Colorimetric Method)<sup>4</sup>

## 3. Terminology

### 3.1 Definitions:

3.1.1 For definitions of terms used in this test method, refer to Terminology D 1356.

### 3.2 Definitions of Terms Specific to This Standard:

3.2.1 *eluent*—the ionic mobile phase used to transport the sample through the ion exchange column.

3.2.2 *resolution*—the ability of a column to separate constituents under specified test conditions.

## 4. Summary of Test Method

### 4.1 Sample Collection:

4.1.1 Air is drawn at a rate of 15 L/min over a continuous 24-h period through three 500-mL glass impingers (in-line) filled with 0.02 *N* sodium bicarbonate [NaHCO<sub>3</sub>] “buffer” solution. A target air volume of 20  $\text{m}^3$  is sampled.

4.1.2 Impinger buffer solution has a pH of 8.2 and was selected to prevent hexavalent chromium from being reduced to trivalent chromium [Cr(III)] in an acidic medium during sampling (4).

4.1.3 The impinger buffer solution from each impinger is analyzed for hexavalent chromium.

### 4.2 Sample Analysis (1, 2, 3, 4)<sup>5</sup>:

4.2.1 A volume of filtered sample, typically 1 mL, is injected into the eluent flow path and separated by anion exchange using an ammonium sulfate [(NH<sub>4</sub>)<sub>2</sub>SO<sub>4</sub>] based eluent.

4.2.2 After separation, the sample is reacted with an acidic solution of diphenylcarbohydrazide. Hexavalent chromium reacts selectively with this reagent to form the characteristic violet colored complex.

4.2.3 The eluent stream passes through a photometric detector for detection of the chromium diphenylcarbohydrazide complex by visible absorbance at 520 nm. Absorbance is proportional to the hexavalent chromium concentration.

<sup>4</sup> Discontinued; See 1991 Annual Book of ASTM Standards, Vol 11.03.

<sup>5</sup> The boldface numbers in parentheses refer to a list of references at the end of the text.

## 5. Significance and Use

5.1 Hexavalent chromium has been shown to be a human respiratory carcinogen in epidemiological studies when humans are exposed to relatively high airborne concentrations. Such high exposures may also induce dermal sensitization to hexavalent chromium in humans (5).

5.2 Ambient atmospheric concentrations of hexavalent chromium are well below detection limits of sampling methods including Test Method D 3586 and NIOSH-7600 (1).

5.3 Objective assessment of ambient atmospheric concentrations of hexavalent chromium provides a means of evaluating exposures to atmospheric hexavalent chromium in a manner that can be related to health-based risk levels. Collecting such actual monitoring data reduces or eliminates the need for theoretical resuspension modeling and provides improved basis for health assessments of potential exposures (5).

5.4 The buffered impinger sampling technique provides pH control of the sampling medium, which stabilizes the oxidation state of hexavalent chromium during sampling (6).

5.5 Ion chromatography provides a means of separating the hexavalent chromium from other species present in the sample, many of which interfere with other detection methods. The combination of this separation with a sensitive colorimetric detection method provides a selective and sensitive analytical method for hexavalent chromium with minimal sample preparation (4).

## 6. Interferences

6.1 Reducing agents may reduce hexavalent chromium to trivalent chromium in acidic matrices. Preservation of a pH 7.8 or greater will minimize the effect of these species. The oxidation of trivalent chromium to hexavalent chromium during this test method is unlikely to occur (6).

6.2 By virtue of the chromatographic separation, essentially all interfering species are removed from the hexavalent chromium before detection. The response of 1 mg/L of hexavalent chromium is not affected by 1000 mg/L of chromic ion.

6.3 Interferences may result from overloading of the analytical separator column capacity with high concentrations of anionic species in the sample. Concentrations of chloride ion or sulfate ion up to the equivalent of 2 % NaCl and 5 % Na<sub>2</sub>SO<sub>4</sub> do not affect the separation or detection when using a 100- $\mu$ L sample loop (2).

6.4 Hypochlorite [OCl<sup>-1</sup>] (100 mg/L) in the buffer solution has been found to cause a positive interference with hexavalent chromium analyses to the extent of 0.3 to 1  $\mu$ g/L. Hypochlorite (1 mg/L) has also been found, in the presence of 50  $\mu$ g/L trivalent chromium, to cause a 1.2- $\mu$ g/L positive interference with hexavalent chromium.

6.5 Permanganate [MnO<sub>4</sub><sup>-1</sup>] (0.5  $\mu$ g/L) causes a positive 0.07- $\mu$ g/L interference with hexavalent chromium.

6.6 No other interferences were observed from 10  $\mu$ g/L BrO<sub>3</sub><sup>-</sup>, MoO<sub>4</sub><sup>-2</sup>, ClO<sub>4</sub><sup>-</sup>, S<sub>2</sub>O<sub>8</sub><sup>-2</sup>, VO<sub>4</sub><sup>-3</sup>, Be<sup>+</sup>, Cu<sup>+2</sup>, Ni<sup>+2</sup>, Ag<sup>+</sup>, Tl<sup>+3</sup>, V<sup>+3</sup>, As<sup>+3</sup>, Ba<sup>+2</sup>, Cd<sup>+2</sup>, Co<sup>+2</sup>, Cr<sup>+3</sup>, Mo<sup>+5</sup>, Sb<sup>+3</sup>, Zn<sup>+2</sup>, Pb<sup>+2</sup>, F<sup>-</sup>, Cl<sup>-</sup>, Br<sup>-</sup>, NO<sub>3</sub><sup>-</sup>, NO<sub>2</sub><sup>-</sup>, P<sub>2</sub>O<sub>6</sub><sup>-4</sup>, SO<sub>4</sub><sup>-2</sup>, 100 mg/L Se, or 1 mg/L Hg (6).

## 7. Apparatus

### 7.1 Sampling Apparatus:

7.1.1 *Impinger Sampling Train*—For a schematic drawing of the major sampling train components see Fig. 1. The sampling train for collecting particulate matter and hexavalent chromium consists of the following elements:

7.1.1.1 *Impingers*—Three 500-mL impingers (in-line) are used in the sampling train. The first two impingers in the series (A and B) use nozzled impinger inlets with impaction plates. These impingers impinge air at high velocity against the impaction plate creating smaller air bubbles which provide more surface area for air contact with buffer solution. The third impinger (C) has a straight inlet nozzle and no impaction plate.

7.1.1.2 *Impinger Buffer Solution*—0.02 N sodium bicarbonate buffer solution (see 8.3.1) is added to the impingers such that: Impinger A = 250 mL, B = 200 mL, and C = 150 mL. These particular impinger sodium bicarbonate solution volumes are recommended to minimize post sample volume disparities between impingers.

7.1.2 The sampling train apparatus is interconnected by the following elements:

7.1.2.1 *Sample Line/Probe*—Sample is drawn from ambient air through a sample line/probe that consists of a 100 to 150-mm polytetrafluoroethylene (PTFE) tube (12-mm (1/2-in.) outside diameter and 9-mm (3/8-in.) inside diameter). The sample line/probe is inserted into the air inlet of the first Impinger (A).

7.1.2.2 Impingers A, B, and C are interconnected using two glass impinger U-joints. The last impinger in the series (C) is connected to the sample pump by means of vinyl tubing using a glass 0.5 $\pi$  radian (90°) angle impinger joint that adapts the




FIG. 1 Diagram of a Sampling Train and Sampling Apparatus

impinger to wax film tubing (see 7.1.3). Impinger clips, wax, and wax film wraps are used to secure all impinger connections and prevent sampling train leaks.

7.1.3 *Sampling Box*—A pre-assembled impinger sampling box holds the impinger sampling train and is designed so that the sample line/probe protrudes outside the box and bends downward. The sample box is fitted with vinyl tubing (14-mm (5/16-in.) outside diameter and 9-mm (3/8-in.) inside diameter) that connects the impinger sampling train to a sample pump (see 7.1.2.2). The vinyl tubing is fitted with an in-line rotameter to facilitate sampling train operational checks.

7.1.3.1 An in-line rotameter fitted on the sample box facilitates operational checks of the sampling system. The rotameter is a glass variable area flow meter capable of measuring flowrates between 10 and 15 L/min, calibrated in accordance with Practice D 3195.

7.1.3.2 *Leakless Sample Pump*—A vane-axial electrically operated sampling pump capable of drawing 10 to 18 L/min of air through the sampling train over 24 h is suitable.

7.1.3.3 *Flow Control Device*—Air flowrate control can be enhanced using a critical orifice or dry gas meter in accordance with Test Methods D 2914. Protect the orifice or gas meter from particulate matter (see 11.2.6).

7.1.4 *Bubble Meter*—The bubble meter is used as a primary method of sampling train air flowrate calibration (see 10.1) and shall be capable of reading sampling air flowrates of 2 to 30 L/min. Connect the bubble meter to the sample line/probe with a flexible rubber tube.

7.1.5 An elapsed time meter is placed in line with the sample pump to assist in detection of electrical interruptions that could have occurred over the 24 h interval.

7.1.6 *Stop Watch or Timer*.

7.1.7 *pH Meter*, to measure the pH of the impinger buffer solution.

7.1.8 *Refrigerator or Ice Cooler*, for storage of samples prior to shipment to the laboratory (see 11.4).

7.1.9 *Ice Cooler*, for transport of samples to the laboratory (see 11.4).

7.1.10 *Meteorological Weather Station or Weather Data Service*, to determine ambient temperature, pressure, relative humidity, wind speed and direction, and precipitation (see 11.2.7). This information may be useful to interpret data, but is not required to correct data for standard conditions.

7.2 *Analytical Apparatus (4)*:

7.2.1 *Ion Chromatograph*—The ion chromatograph shall have the following components as shown in Fig. 2.

7.2.1.1 *Pump*, capable of delivering a constant flow in the range of 1 to 5 mL/min at a pressure of 15 to 150 MPa (200 to 2000 lb/in.<sup>2</sup>).

7.2.1.2 *Injection Valve*—A low dead-volume valve that will allow the loading of a sample contents into the eluent stream. Sample loops of up to 1 mL will provide enhanced detection limits. Smaller sample loops will result in proportionally higher detection limits.




FIG. 2 Diagram of an Ion Chromatograph Using Post-Column Reagent Addition and Photometric Detection

7.2.1.3 *Guard Column*—A column placed before the separator column to protect the separator column from fouling by particles or strongly absorbed organic constituents.

7.2.1.4 *Separator Column*—A column packed with high capacity pellicular anion exchange resin that is suitable for resolving hexavalent chromium from a sample containing high total dissolved solids (for example, 3 % Na<sub>2</sub>SO<sub>4</sub>).

7.2.1.5 *Reagent Delivery Module*—A device capable of delivering 0 to 2 mL/min of reagent against a backpressure of up to 40 kPa (6.0 lb/in.<sup>2</sup>).

7.2.1.6 *Mixing Tee and Reaction Coil*—A device capable of mixing two flowing streams with minimal band spreading.

7.2.1.7 *Detector*—A low-volume, flow-through visible absorbance detector with a nonmetallic 1-cm flow path. The detection wavelength for hexavalent chromium is 520 nm.

7.2.1.8 *Recorder, Integrator, or Computer*—A device compatible with detector output, capable of recording detector response as a function of time for the purpose of measuring peak height or area.

7.2.2 *Eluent Reservoir*—A container suitable for storing eluent.

7.2.3 *0.45 μm syringe filter*, for sample filtration prior to analysis (see 11.5.7).

7.2.4 *Syringe*—A syringe equipped with a male fitting and a capacity of at least 1 mL or auto sampler module (see 11.5.8).

## 8. Reagents and Materials

8.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. All reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.<sup>6</sup>

8.2 *Purity of Water*—Water shall be Type II reagent water conforming to Specification D 1193.

### 8.3 *Sampling Reagents and Materials:*

8.3.1 *Impinger Buffer Solution*—0.02 N sodium bicarbonate buffer solution: dissolve 1.67 g of sodium bicarbonate ( $\text{NaHCO}_3$ ) in 1 L of reagent water.

8.3.2 *Impinger Buffer Solution Spike*—Prepared in 0.5, 1, and 10- $\mu\text{g/L}$  concentrations by diluting appropriate volumes of the 1000  $\mu\text{g/L}$  hexavalent chromium standard (see 8.2.2) in the buffer solution (see 8.3.1).

8.3.3 *1 % Nitric Acid Wash Solution*—Dilute 10 mL of concentrated reagent grade nitric ( $\text{HNO}_3$ ) acid, sp gr 1.42, to 1 L with water.

### 8.4 *Ion Chromatography Eluents:*

8.4.1 *Eluent Concentrate* (2.0 M  $(\text{NH}_4)_2\text{SO}_4$ , 1.0 M  $\text{NH}_4\text{OH}$ )—Dissolve 264 g of ammonium sulfate  $(\text{NH}_4)_2\text{SO}_4$  in about 500 mL of water. Add 65 mL of concentrated ammonium hydroxide ( $\text{NH}_4\text{OH}$ —sp gr 0.90). Mix well and dilute to 1 L in a volumetric flask.

8.4.2 *Eluent* (0.20 M  $(\text{NH}_4)_2\text{SO}_4$ , 0.1 M  $\text{NH}_4\text{OH}$ )—Add 100 mL of eluent concentrate (see 8.4.1) to a 1-L flask and dilute to volume with water.

8.4.3 *Diphenylcarbohydrazide Reagent*—Dissolve 0.5 g of 1,5-diphenylcarbohydrazide in 100 mL of reagent grade methanol. Add to about 500 mL of water containing 28 mL of 96 % sulfuric acid (sp gr 1.84). Dilute with water to 1 L in a volumetric flask.

### 8.5 *Calibration Standards:*

8.5.1 *Hexavalent Chromium Solution, Stock* (1000 mg Cr/L)—Dissolve 0.2828 g of potassium dichromate ( $\text{K}_2\text{Cr}_2\text{O}_7$ ) that has been dried at 105°C for 1 h, in water. Dilute to 100 mL in a volumetric flask.

8.5.2 *Hexavalent Chromium Solution, Standard* (1000  $\mu\text{g}$  Cr/L)—Pipet 1.00 mL of the chromium stock solution (see 8.5.1) into a 1-L volumetric flask and dilute to volume with water.

8.5.3 *Hexavalent Chromium Solution, Calibration Standards*—Standards are prepared in 5, 1, 0.5, 0.05, and 0.02- $\mu\text{g/L}$  concentrations by diluting appropriate volumes of the 1000- $\mu\text{g/L}$  standard in the impinger buffer solution (see 8.3.1).

## 9. Sampling

9.1 Select a sampling location to provide information on the possible impact of site activities, conditions, and possible human exposures. Collect both upwind and downwind ambient air samples.

9.2 To assess ambient environmental concentrations of hexavalent chromium, collect samples with a target air volume of 20  $\text{m}^3$  over a continuous 24-h sampling interval.

### 9.3 *Field Quality Assurance and Control Samples (QA/QC):*

9.3.1 Field QA/QC samples collected include: one impinger field blank for every sampling period (see 11.1.2.4 and 12.3.2).

9.4 Sampling collection and analytical procedures are described in Section 11.

9.5 For general information on sampling refer to Practice D 1357.

## 10. Calibration and Standardization

### 10.1 *Sampling Calibration:*

10.1.1 Calibrate sample air flowrate using a primary method of calibration at the beginning (pre-calibration) and end (post-calibration) of each sampling session as follows. The beginning and end air flowrates shall not vary by more than  $\pm 30\%$  (see 11.2.6). The final sample flowrate is an average of the pre- and post-sample calibrations (7). Perform maintenance and repairs to calibration equipment in accordance with the manufacturer's instructions, and keep records for documentation.

10.1.2 Use a soap bubble meter (see 7.1.4), rotameter (see 7.1.3.1), or both, for sample calibration. Procedures for sample calibration using a soap bubble meter are provided in 10.1.3 to 10.1.8.

10.1.3 Wear latex gloves during sample calibration procedures to prevent sample contamination.

10.1.4 Activate the sample pump (see 7.1.3.2) and allow it to stabilize. After the pump flowrate stabilizes, attach the sampling train to the pump by means of the vinyl tubing (see 7.1.3).

10.1.5 Check the sample train for leaks by pressurizing the sample impinger train. This can be accomplished by restricting flow through the train at the air inlet, while being careful not to draw buffer solution into the sample pump or from Impinger A to B, or B to C. When the impinger train is pressurized, inspect each impinger for signs of bubbling. If bubbling is observed, inspect the connections (see 7.1.2.1 and 7.1.2.1) for leaks. These connecting joints may require additional wax or wax film wrapping, or both, (see 7.1.2.2). If bubbling does not cease, replace the sampling train with clean decontaminated apparatus.

10.1.6 Attach the bubble meter to the air line/probe of the sampling train, and prime the bubble meter flow cell with bubble solution by drawing repeated films through the cell until a single film travels the distance.

10.1.7 Adjust the sample pump so that the sample flowrate is 15 L/min ( $\pm 30\%$ ). Record the average of at least three calibration runs in a sample log book. Note the time and date, and set the elapsed timer to zero.

### 10.2 *Analytical Calibration:*

10.2.1 Prepare hexavalent chromium solution standards as described in 8.5.3.

10.2.2 Determine the analytical instrument chromium response for each of the standards using the procedure defined in 11.5.

<sup>6</sup> *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

10.2.3 Prepare a calibration curve by using a linear plot of the peak height or area as a function of standard concentration by the regression analysis of least squares. The coefficient of determination ( $R^2$ ) shall be greater than 0.99.

10.2.4 Use the same procedure to determine sample results as was used to prepare the calibration curve; for example, peak height or peak area as a function of concentration (see 10.2.3).

10.2.5 Prepare a new calibration curve when new reagents are made, hardware is altered, or continuing calibration varies from the initial calibration by more than 10 %.

## 11. Procedure

### 11.1 *Sampling Train Assembly:*

11.1.1 *Decontamination*—Decontaminate the sampling train impingers, impinger connecting U-joints, sample line/probe, and the graduated cylinders used for measuring volumes of buffer solution in a 1 % nitric acid wash (see 8.3.3) prior to assembly. To ensure complete decontamination, the following decontaminating procedure is recommended. Immerse all components to be decontaminated in a 1 % nitric acid wash (see 8.3.3) for 10 min. The 1 % nitric acid immersion is followed by two consecutive reagent water immersion washes for 10 min each. Rinse three times with reagent water and air dry the apparatus, then cover all openings securely with wax film to prevent glassware contamination.

11.1.2 *Sampling Train Pre-Assembly*—Assemble the sampling train as follows using latex gloves to prevent contamination:

11.1.2.1 Measure pH of the impinger buffer solution (see 8.3.1) and record the reading and buffer solution lot number. The pH shall be greater than 7.8. If the pH is less than 7.8, discard the solution.

11.1.2.2 Remove wax film (that prevents glassware contamination, see 11.1.1) from glassware openings. Rinse sampling train apparatus (see 7.1.2): impinger, impinger connectors, graduated cylinders, and polytetrafluoroethylene (PTFE) tubing with impinger buffer solution (see 8.3.1).

11.1.2.3 Using a graduated cylinder, add 250, 200, and 150 mL of impinger buffer solution to Impingers A, B, and C, respectively (see 7.1.2.2).

11.1.2.4 Wrap neck of impinger (where impinger barrel and impinger are joined) with wax film to prevent sample train leaks.

11.1.2.5 Prepare one impinger field blank (see 9.3.1) as described in 11.1.1 through 11.1.2.4 with 200 mL of impinger solution (see 8.3.1). Cover the openings with wax film to prevent contamination with atmospheric hexavalent chromium, and place the field blank into a sampling box with a sampling train and allow it to remain there for the duration of sampling. Recover the field blank impinger sample for analysis as described in 11.3.

### 11.1.3 *Sampling Train Assembly:*

11.1.3.1 Place Impingers A, B, and C (see 7.1.2.2) in order within the sampling box.

11.1.3.2 Insert the sample line/probe (see 7.1.2.1) into air inlet of Impinger A and wrap connecting joint with wax film.

11.1.3.3 Using glass impinger U-joint connectors, connect Impinger A air exhaust opening to air intake of Impinger B, and connect Impinger B and C in the same manner. Seal the point

of contact between each U-joint and impinger with wax film (see 8.3.3) and secure it with an impinger clip.

11.1.3.4 Fasten a  $0.5\pi$  radian glass connector to the air exhaust of Impinger C. Wrap this impinger joint with wax film and secure with a clip. Connect this sample outlet to a sample pump using vinyl tubing (see 7.1.3).

### 11.2 *Sampling Operation:*

11.2.1 Select a sampling area as described in 9.1.

11.2.2 Calibrate the sampling air flowrate as described in 10.1.1 through 10.1.7.

11.2.3 Record the readings on the elapsed time meter and the dry meter, if used. Record the precise time and date that the sampling is started. Record the starting reading of the rotameter. Sample air through sample apparatus for 20 to 24 h.

11.2.4 Perform sampling equipment checks to verify buffer solution volumes and air flow rotameter readings, and note evidence of sample disruption.

11.2.5 Before stopping the sampling, record the measurements listed in 11.2.3. Post-calibrate the sample air flow as described in 10.1.1 through 10.1.7, and perform a leak test as described in 10.1.5.

11.2.6 If the final flowrate varies more than  $\pm 30$  % from the initial reading, label the flow data for the sample as “suspect.” Inspect the sample pump and sample train to determine the cause of the flowrate deviation (see 7.1.3.3).

11.2.7 Record ambient weather conditions during sampling, for example, temperature, pressure, relative humidity, precipitation, and wind speed and direction, if taken (see 7.1.10 and 12.3.3.1).

### 11.3 *Sample Recovery:*

11.3.1 Use a separate clean (decontaminated) graduated cylinder for each impinger sample buffer solution collected.

11.3.2 Wear latex gloves during sample recovery to prevent contamination. Change gloves frequently to avoid cross-contamination between samples.

11.3.3 Remove sample line/probe (see 7.1.2.1) from Impinger A, rinse inside line/probe into the graduated cylinder for Impinger A. Rinse inside each Impinger A opening, and remove and rinse the impinger from the impinger barrel. Pour the impinger sample buffer solution from Impinger A into the graduated cylinder for Sample A.

11.3.4 Remove and rinse inside the U-joint connecting Impinger A and B into the graduated cylinder for B. Rinse inside each Impinger B opening, and remove and rinse the impinger from the impinger barrel. Pour the impinger sample buffer solution from Impinger B into the graduated cylinder for Sample B.

11.3.5 Remove and rinse inside the U-joint connecting Impinger B and C into the graduated cylinder for Impinger C. Rinse inside each impinger opening, and remove and rinse the impinger from the impinger barrel. Pour the sample impinger buffer solution from the impinger barrel into the graduated cylinder for Sample C.

11.3.6 Record volume (mL) of impinger buffer solution and rinsate collected from each impinger.

11.3.7 Place each impinger sample solution in a “lab-clean” sample bottle.

11.3.8 Measure pH of each impinger sample solution and record reading. (Rinse the pH probe thoroughly with reagent water after each use.)

11.3.9 The sample pH shall be greater than 7.8. If the pH is less than 7.8, discard the sample.

11.3.10 Label the sample jar with the following information: sample identification number, date of sample collection, laboratory analysis requested, impinger buffer solution volume, and final sample pH.

11.4 *Sample Handling*—To retard chemical reactivity of hexavalent chromium, refrigerate the samples until shipment to the laboratory (see 7.1.8). Place the samples in iced coolers for laboratory shipment (see 7.1.9). Upon receipt in the laboratory, store the samples at 4°C until analyzed. For information on hexavalent chromium stability during sample storage, refer to 13.3.

11.5 *Hexavalent Chromium Analysis (3, 4):*

11.5.1 Verify the sample buffer solution volumes and pH upon receipt of samples from the field (see 11.3.9).

11.5.2 Set-up the ion chromatograph (see 7.2.1) in accordance with the manufacturer’s instruction.

11.5.3 Install organic guard column (see 7.2.1.3) and separator columns (see 7.2.1.4) in the ion chromatograph.

11.5.4 Install a 1 mL sample loop on the injection valve (see 7.2.1.2) of the ion chromatograph.

11.5.5 Adjust the eluent (see 8.4.2) flowrate to 1.5 mL/min. Increase the flow of the diphenylcarbohydrazide reagent (see 8.4.3) until the flowrate is 2.0 mL/min. Measure the pH of the detector effluent to confirm it is two or lower.

11.5.6 After the flowrates are adjusted, allow the system to equilibrate for about 15 min.

11.5.7 Stir the samples, remove a portion, and filter it through a 0.45 µm syringe filter (see 7.2.3).

11.5.8 Inject 1 mL of filtered sample through the sample port using an appropriate syringe or auto sampler (see 7.2.4), into the eluent stream and mark the injection time on the chromatogram recorder (see Fig. 3).

12. Calculation

12.1 *Sampling (7):*

12.1.1 Calculate the sample air volume ( $V_s$ ), in  $m^3$  using the sample time and average flowrate as follows:

$$V_s = [(F_s + F_e) / 2] [T_s \times 1000]$$

where:

$F_s$  = starting flowrate (L/min),

$F_e$  = ending flowrate (L/min), and

$T_s$  = total sampling time, min.

12.1.2 The final sample air flowrate is determined by averaging the pre- and post-sample air flowrates in  $m^3$  (see 7.1.10 and 10.1.1).

12.1.3 Record the final sample air volume of air sampled in  $m^3$ .

12.2 *Analytical (4):*

12.2.1 Determine the hexavalent chromium concentration in µg/L, using the same method that was used in the calibration step, that is, peak height or area from the calibrated curve (see 10.2.3).




FIG. 3 Ion Chromatographic Determination of Hexavalent Chromium

12.2.2 For samples that have been diluted, calculate the original hexavalent chromium concentration in µg/L by the following:

$$Cr(VI) \mu\text{g/L} = (C \times F) / V$$

where:

$C$  =  $Cr(VI)$  µg/L, read from the calibration curve (see 10.2.3),

$F$  = volume of diluted sample, in mL, and

$V$  = volume of undiluted sample in mL.

12.3 *Sampling Results Reporting:*

12.3.1 After analytical determination of the hexavalent chromium concentration from each impinger in the sampling train, determine the net hexavalent chromium quantity by subtracting the field blank (see 9.3.1) value from all impinger samples with detectable amounts of hexavalent chromium. If

the field blank is below the analytical method limit of detection, then no corrections are made to sample values. Note the field blank values on the sampling report.

12.3.2 Determine the hexavalent chromium sample concentration by totaling the reported quantities from each impinger (A, B, C) in a given sample train according to the following:

12.3.2.1 If an impinger sample contains hexavalent chromium below the analytical method limit of detection, then its value is not included in the total.

12.3.2.2 If all three impingers are below the analytical method limit of detection, the value for Impinger A (the first impinger in the sampling train series) is reported.

12.3.3 Report the hexavalent chromium concentration in terms of  $\mu\text{g}/\text{m}^3$  or  $\text{ng}/\text{m}^3$  of air sampled.

12.3.3.1 Report the meteorological data if collected. If this information was not recorded, state this fact to complete the report.

### 13. Precision and Bias (8)

13.1 *Precision*—The precision of this test method has been tested using replicate sample tests. In these sampling tests, the replicate sampling apparatus as described in 7.1 having sampling lines (see 7.1.2.1) arranged a minimum of 3 ft from each other were operated. The pooled coefficient of variation (CV) between three groups of three replicate sample tests ranged from 3.1 to 20 %. The CV of 20 % was for sample concentrations within a factor of five of the method limit of detection. Four additional replicate sample tests were conducted on other sites. The pooled coefficient of variation for these replicate tests was 10.8 %. These results are within the range specified by the USEPA-Contract Laboratory Program (CLP) for quantification of trace metals (8).

13.2 *Bias*—There is no hexavalent chromium particulate reference standard; therefore, generation of a known atmospheric concentration in a test chamber is not a feasible method to determine method bias. Test method accuracy was assessed as a function of spike recovery by substituting 1  $\mu\text{g}/\text{L}$  hexavalent chromium buffer solution spike (see 8.3.2) for the buffer solution (see 8.3.1) in the test procedure as described in Section 11. In these tests, two co-located groups of three replicate 1  $\mu\text{g}/\text{L}$  spike tests were operated. After the spiked sample tests were corrected for background concentrations of hexavalent chromium, spike recovery ranged from 87 to 101 %, and the mean recovery was 94 %. In addition, there was no statistical significance between the two groups of three replicate spike tests at the 95 % confidence limit. Other site independent evaluations of replicate 10  $\mu\text{g}/\text{L}$  spike tests showed a corrected spike recovery range of 90 to 118 %. These results showed that method accuracy is within the USEPA-CLP for recovery criteria (8).

13.3 The bias of this test method cannot be checked against other methods since the unit in which it reports is defined by the test method. As an indication of test method hexavalent chromium collection and stabilization, test method results were

compared with co-located total chromium air sampling results. Ambient air studies have shown hexavalent chromium to be 10 to 40 % of the total chromium in the vicinity of hexavalent chromium sources (6). In 77 co-located tests, results showed that the average percentage of hexavalent chromium in the atmospheric concentration of total chromium was 25 % (8).

13.4 Spike recovery tests as described in 13.2 and 13.3 also indicate that hexavalent chromium is stable in this test method and is not reduced to trivalent chromium. Additional laboratory and test method spike studies have shown the stability of hexavalent chromium in the buffer solution as a function of storage time. In a laboratory study, a 10  $\mu\text{g}/\text{L}$  spike buffer solution sample was analyzed on days 2, 6, 8, 10, 13, 15, and 20 days after spiking. The relative percent difference (RPD) in hexavalent chromium concentration over time ranged from  $-4.9$  to  $8.3$  %, with a mean RPD of  $-0.4$  %. These results show that hexavalent chromium is stable for up to 20 days when stored at  $4^\circ\text{C}$ . In a field test method study, five 10  $\mu\text{g}/\text{L}$  spike samples were analyzed over a period of 104 days. The initial date of analysis was assigned a time of zero and subsequent analyses were performed at days 24, 53, 75, 87, and 104. Only small reductions ( $<8$  %) were noted when these samples were stored for up to 104 days at  $4^\circ\text{C}$ . Method tests have also been performed to evaluate the conversion of trivalent chromium to hexavalent chromium. Results indicated that a small fraction of trivalent chromium may be converted to hexavalent chromium, but the conversion is too small to have a measurable effect on hexavalent chromium concentrations of total airborne chromium typically found in the urban air (8).

13.5 The sample train air flowrate has been evaluated by co-located samples collected at flowrates at 5, 10, and 15 L/min of air. No statistical difference was found between first impinger results using the *F*-Test ( $\alpha = 0.05$ ). Comparisons made between samples collected at 30 L/min (1  $\text{ft}^3/\text{min}$ ) and 15 L/min (0.5  $\text{ft}^3/\text{min}$ ) indicated no increased collection efficiency at the higher flowrate, however, solution agitation and carry over between impingers were greater at the 30 L/min flowrate.

13.6 The analytical method limit of detection (LOD) was assessed using the guidelines published by the Environmental Protection Agency, in Test Methods for Evaluating Solid Waste (9) for assessing method detection limit. Using SW846 procedures, the LOD of this test method was calculated to be 0.018  $\mu\text{g}/\text{L}$ . The practical limit of quantification (PQL) was evaluated at ten times the standard deviation of replicate results, as is consistent with the definition published by the American Chemical Society. The PQL was calculated to be 0.065  $\mu\text{g}/\text{L}$ . Analytical response was linear from the range of 0.02 to 5.0  $\text{ng}/\text{mL}$  (see 10.2.2) with a correlation coefficient greater than 0.999.

### 14. Keywords

14.1 ambient atmospheres; atmospheres; chromium; hexavalent chromium; sampling

**REFERENCES**

- (1) Molina, D., and Abell, M. T., *NIOSH Manual of Analytical Methods*, 3rd. ed., Method No. 7600, National Institute of Occupational Safety and Health (NIOSH), Pub. No. 84-100, Cincinnati, OH, 1984.
- (2) U.S. Environmental Protection Agency (USEPA), "Method 218.6 Determination of Dissolved Hexavalent Chromium in Drinking Water, Groundwater, and in Industrial Waste Water Effects by Ion Chromatography," Environmental Monitoring Systems Laboratory, Office of Research and Development, Cincinnati, OH, 1990.
- (3) U.S. Environmental Protection Agency (USEPA), "Method 7196 Hexavalent Chromium: Colorimetric Method," Environmental Monitoring and Support Laboratory, Cincinnati, OH, 1984.
- (4) Dionex Technical Note, *Determination of CrVI in Water, Wastewater, and Solid Waste Extracts*, TN26, May 1990.
- (5) Paustenbach, D. J., W. E. Rinehart, and P. J. Sheehan, "The Health Hazards Posed by Chromium-Contaminated Soils in Residential and Industrial Areas: Conclusions of an Expert Panel," *Reg Toxicol. Pharmacol.* 13:195–222, 1991.
- (6) Research Triangle Institute (RTI), *The Fate of Hexavalent Chromium in the Atmosphere*, RTI/3789/00-01F, Air Resources Board, 1988.
- (7) American Conference of Governmental Hygienists, *Air Sampling Instruments for Evaluation of Atmospheric Contaminants*, 5th Edition, Section I, "Calibration of Air Sampling Instruments," Cincinnati, OH, 1978.
- (8) Sheehan, P. J., Ricks, Ripple, and Paustinbach, D. J., "Field Evaluation of Sampling and Analytical Method for Airborn Hexavalent Chromium," *American Industrial Hygiens Association Journal*, January, 1992.
- (9) U.S. Environmental Protection Agency (USEPA), *USEPA Test Methods for Evaluating Solid Waste, (SW 846)*, Revision 1987, 1983.

*ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.*

*This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.*

*This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).*