



Standard Practice for Detection of Mycoplasma Contamination of Bovine Serum by the Large Volume Method¹

This standard is issued under the fixed designation E 1536; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice covers the procedures used for detection of mycoplasma contamination in serum by direct microbiological culture.

1.2 This practice does not cover procedures used for detection of mycoplasma in cell cultures.

1.3 This practice does not cover indirect methods for detection of mycoplasma contamination.

1.4 This practice does not cover methods for identification of mycoplasma cultures.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:

E 1531 Practice for the Detection of Mycoplasma Contamination of Cell Cultures by Growth on Agarose Medium²

E 1532 Practice for the Detection of Mycoplasma Contamination of Cell Cultures by the Use of the Bisbenzamide DNA-Binding Fluorochrome²

E 1533 Practice for Indirect Detection of Mycoplasma in Cell Culture by 4'-6-Diamidino-2-2 Phenylindole (DAPI) Staining²

3. Terminology

3.1 Definitions:

3.1.1 *direct mycoplasma detection, n*—demonstration of characteristic colonial growth on axenic agar medium.

3.1.2 *large volume testing, n*—using a large volume inoculum in an enrichment culture.

3.1.3 *mycoplasma (Mollicute), n*—smallest prokaryotes capable of self replication.

4. Significance and Use

4.1 Mycoplasmas of bovine origin are prevalent contami-

nants of cell cultures. Contamination can be detected by the large volume method.^{3,4}

4.2 Heat inactivated serum need not be tested for mycoplasmas. Heating serum to 56°C for 30 minutes will kill mycoplasmas.

4.3 Mycoplasmas may be present in any particular lot of serum but may not be detected because of inadequate sample size; thus, negative test results do not provide absolute assurance that the test serum is free of mycoplasmas.

5. Liquid Medium Preparation

5.1 Add 105-g mycoplasma broth base, 5-g glucose, 5-g arginine, and 20 mL of a 0.5 % solution of phenol red to 4080 mL of distilled water. Mix to dissolve ingredients.

5.2 Dispense medium, in 400-mL amounts into 500-mL screw-capped bottles.

5.3 Autoclave.

5.4 Sterile refrigerated medium is stable for four months.

6. Quality Control

6.1 Prior to testing large volumes of bovine serum, check sterility and ability of liquid medium to support mycoplasma growth.

6.2 Strains used to test for growth support: *M. arginini*, G230, *M. bovis*, Donetta; *A. laidlawii*, PG8.

6.3 For quality control, a portion of the base liquid medium is supplemented with 20 % of newborn calf serum. This batch of serum must be extensively tested to ensure that it is free of mycoplasma contamination and it should be in sufficient quantity to last for an extended period of time. Challenge mycoplasma strains for the quality control test should be diluted so that approximately 100 colony-forming units are contained in the inoculum.

7. Test Procedure

7.1 The sample is 100 mL of uninactivated bovine or equine serum. Multiple samples will increase the probability of mycoplasma detection.

¹ This practice is under the jurisdiction of ASTM Committee E-48 on Biotechnology and is the direct responsibility of Subcommittee E48.02 on Characterization and Identification of Biological Systems.

Current edition approved May 10, 2000. Published June 2000. Originally published as E 1536 – 93. Last previous edition E 1536 – 93.

² *Annual Book of ASTM Standards*, Vol 11.05.

³ Barile, M. F., and Kern, J., "Isolation of *Mycoplasma arginini* from Commercial Bovine Sera and Its Implication in Contaminated Cell Cultures," *Proceeding of the Society for Experimental Biology and Medicine*, 138, 1971, pp. 432–437.

⁴ Del Giudice, R. A., Tully, J. G., "Isolation of Mycoplasmas from Cell Cultures by Axenic Cultivation Techniques," *Molecular and Diagnostic Procedures in Mycoplasmaology*, Joseph G. Tully and Schmucl Razin, Eds., Academic Press, 1996, Vol II, pp. 411–418.

7.2 Inoculate one 100-mL sample of fetal bovine serum into 400 mL of medium, and incubate for 21 days at 37°C.

7.3 After incubation for 5, 10, and 21 days, 0.1 mL is subcultured to each of two agar plates (see Practice E 1531). Incubate one plate anaerobically and one plate aerobically. Examine all plates after incubation for 5 and 14 days.

7.4 Serum is considered contaminated if typical mycoplasma colonies grow on the agar medium.

8. Keywords

8.1 mycoplasma; serum

The American Society for Testing and Materials takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).